Fermatix
360 Audubon Road
Wakefield MA 01880
www.fermatix.com

CERTIFICATE OF ANALYSIS

Product name: tRNA – *E. coli*
Product number: N0101
Lot Number: N0101AA1
Expiration Date: 09-Jul-27
Storage: -20°C for 24 months

Size: ≥20 mg when resuspended

Quality Control Analysis

Parameter	Specification	Result
A260/280 ratio	> 1.7	2.1
RNase detected	No	No
Micro RNA	≤15%	1.3%
Correct electrophoretic profile	Yes	Conforms



Manufactured by Fermatix in compliance with ISO 9001:2015 and ISO 13485:2016 certified Quality Management Systems.

Authorized Quality Control Signature

Jessica R. Kenneson / Senior Biochemist

Authorized Quality Assurance Signature

Judith Harrington / Quality Assurance Manager